


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-5B-PW-PS-N-RIA*

S.N.: *5528*

Date: *JUNE 7 1990*



- a) Output signal amplitude:
0 TO -500 V_{TO}
- b) Pulse width: *R_L > 10 K*
0.1 TO 100 μS_{EC}
- c) Rise time:
≤ 20 NS_{EC}
- d) Fall time:
≤ 20 NS_{EC}
- e) PRF:
PW ≤ 5 μS_{EC} : 0 TO 1 KHz
PW = 100 μS_{EC} : 0 TO 50 Hz
- f) Jitter, stability:
OK
- g) Prime power:

50 VOLTS/DIV
R_L = 10 K...

120/240 V
50 ~ 60 Hz