


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ARR-3-PW-PS-P-RIA*
S.N.: *5529*
Date: *JUNE 7 1990*



50 VOLTS/DIV
R_L = 50 Ω
PRF = 50 kHz

- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: *(TO 50 ns)*
- c) Rise time:
0.1 TO 100 nSEC
 ≤ 10 nSEC
- d) Fall time:
 ≤ 10 nSEC
- e) PRF:
PW ≤ 5.0 nSEC; 0 TO 1 kHz
PW = 100 nSEC; 0 TO 50 kHz
- f) Jitter, stability:

OK
g) Prime power:
120 / 240 V
50 - 60 kHz

[Signature]