


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-15-C-PN-0T*


S.N.: *5531*

Date: *MAY 24 1990*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V.
50-60 Hz.

POUT 40 db ATTEN
5 VOLTS/DIV
(RISE TIME CHECK)



POUT 40db ATTEN
5 VOLTS/DIV
PRF = 100 kHz

[Signature]