


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-AN-1-PS-05-P*

S.N.: *5535*

Date: *JULY 20 1990*



*60 db ATTEN*  
*100 VOLTS DIV*  
*PRF = 5.0 KHz*

- a) Output signal amplitude:  
*0 TO +100 VOLTS*
- b) Pulse width:  
*3 TO 100 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 to 100 Hz*