


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3A-C-P*
S.N.: *5536*
Date: *July 17 1990*



- a) Output signal amplitude:
25 TO 75 AMP
- b) Pulse width:
2.5 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
~ 100 VOLTS/DIV
R_L = 4 Ω
~ 25 AMP/DIV
PRF = 20 KHZ

120/240 V
50 - 60 MHz

[Signature]