


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVA-5-7

S.N.: 5540

Date: June 4 1990



40 db ATTEN

10 VOLT/DIV

PRF \approx 200 KHz

a) Output signal amplitude:

0 TO +20 VOLTS

b) Pulse width:

\leq 500 PSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 V

150 mA

(MAX)

[Signature]