


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MMH-1-P*
S.N.: *5541*
Date: *JUNE 4 1990*



40 db ATTN
10 VOLT/DIV
PRFM 7.0 MHz

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 200 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VOLTS
400 mA (max)

[Signature]