


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-C-PN-05-M*
S.N.: *5550*
Date: *JUNE 4 1990*



50 db ATTEN
~.320 VOLTS/DIV
PRF ~ 1.0 kHz

- a) Output signal amplitude:
0 TO ±350 V
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 2 nSEC
- d) Fall time:
≤ 3 nSEC
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz