


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AV4-S-1-C-P*
 S.N.: *5552*
 Date: *JULY 23 1990*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 135 PSEC.
- c) Rise time:
NA
- d) Fall time:
NA

① *40 db ATTEN : 10 000/DIV*
PRF = 100 KHz

- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK



- g) Prime power:
120/240 V
50 to 60 Hz.

[Handwritten signature]

② *As ① BUT 200 PSEC/DIV*