


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMH-5-1-C-P-05*

S.N.: *5553*

Date: *Aug 3 1990*



- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *≤ 130 PSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

*40 db ATTEN*  
*10 VOLTS/DIV*  
*PRF ≈ 100 KHz*

*50-60 Hz*