



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMMP-1A-C-P-PN-M*
 S.N.: *5554*
 Date: *JULY 27 1990*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-80 Hz

(A) 40 db ATTEN
5 VOLTS/DIV
PRF ≈ 100 KHz



(B) AS (A) BUT 1.0 NSEC/DIV
(RISE TIME)

120/240V
50-80 Hz
[Signature]