


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-P*
S.N.: *5558*
Date: *JUNE 27 1990*



50 dB ATT
32 VOLTS/DIV
PRF = 50 KHZ

- a) Output signal amplitude:
0 TO 40 VOLTS
- b) Pulse width:
2 TO 100 NSEC
- c) Rise time:
≤ 350 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
DW MIN: 0 TO 1 MHz
DW MAX: 0 TO 20 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+5 VOLTS,
150 mA (MAX)