


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-130-UCLA2-PS

S.N.: 5559

Date: JAN 24 1990



a) Output signal amplitude:
≈ +35 VOLTS

b) Pulse width:
50 NSEC

c) Rise time:
≤ 500 PSEC


d) Fall time:
2 NSEC

e) PRF:
0 TO 20 KHz


f) Jitter, stability:
OK

g) Prime power:
120/200 V
50-60 MHz

Ⓐ INPUT 5KHZ CLK
(TYPICAL)



Ⓑ OUTPUT 50 dB ATTEN



Ⓒ OUTPUT OF AVR-SP-4A-UCLA
(40 dB ATTEN)
10 VOLTS/DIV