



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-1A3-P*
 S.N.: *5562*
 Date: *SEPT 7 1990*



Ⓐ *60 db ATTEN*
100 VOLTS/DIV
PRF = 5.0 KHz



Ⓑ *AS Ⓐ BUT*
1.0 NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 4 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
350 mA.
(MAX)