


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-5-1-C-P*

S.N.: *5566*

Date: *July 31 1990*



*40 dB ATTEN*  
*1.0 VOUTS / DIV*  
*PRF ≈ 100 kHz*

- a) Output signal amplitude:  
*0 TO +10 VOUTS*
- b) Pulse width:  
*≈ 130 psec*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 kHz*

*[Signature]*