


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-B1-C-W-P*


S.N.: *5570*

Date: *AUG 29 90*



- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
0.1 TO 5.0 USEC
- c) Rise time:
≤ 5 NSEC
- d) Fall time:
≤ 5 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 60 Hz

① *R_L = 50 Ω*
0.5 USEC/DIV
10 VOLTS/DIV



② *R_L = 50 Ω*
40 dB ATTEN
10 VOLTS/DIV
10 NSEC/DIV
(RISE TIME)

[Signature]