


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-7*
S.N.: *5571*
Date: *AUG 8 1990*



- a) Output signal amplitude:
> 30 VOLTS
- b) Pulse width:
≤ 1 nSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

50 Ω ATTEN
32 VOLTS/DIV
PRF = 500 kHz

+15 ± 1.0 VDC
100 mA (MAX)
[Signature]