


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPB-2-C-P-PN*


S.N.: *5575*

Date: *OCT 12 1990*



- a) Output signal amplitude:
0 TO ±20 VOLTS
- b) Pulse width:
A) 0.4 TO 7 NS
B) 7 NS TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 300 PS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 160 W

*ⓐ 40db ATTEN
10 VOLTS/DIV
A RANGE*



*ⓑ 40 db ATTEN
10 VOLTS/DIV
B RANGE*

[Signature]