

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


APD-5B-C-PN

S.N.:

5576

Date:

OCT 12 1990



- a) Output signal amplitude:  
0 TO  $\pm 40$  AMPS TO
- b) Pulse width: 0.75  $\mu$ s  
25 TO 500 NS
- c) Rise time:  
 $\leq 10$  NS
- d) Fall time:  
 $\leq 10$  NS
- e) PRF:  
0 TO 1.0 KHz
- f) Jitter, stability:

OK

- g) Prime power:

120 / 240 V  
50-60 Hz

10 VOLTS / DIV.  
R<sub>LOAD</sub> = 0.7  $\Omega$