


PULSE GENERATOR

PERFORMANCE CHECK

Model: - AVI-V-HV2-P-CED

S.N.: 5582

Date: SEPT 10 1990



60 db atten
100 VOLT/DIV
RISK = 20 KHZ

- a) Output signal amplitude:
0 TO 1100 VOLTS
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 \pm 1.0 VOLTS
300 mA (MAX)