


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PI-V-HV2-D-CEA*

S.N.: *5583*

Date: *SEPT 10 1990*



60 dB ATTEN

100 VOLTS DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO 100 VOLTS
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT

*300 mA
(MAX)*

[Signature]