


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-C-N

S.N.: 5590

Date: SEPT 19 1990



- a) Output signal amplitude:  
0 TO -350. ✓
- b) Pulse width:  
5 TO 100 NS
- c) Rise time:  
≤ 2 NS
- d) Fall time:  
≤ 3 NS
- e) PRF:  
0 TO 5 KHz

f) Jitter, stability:  
OK

g) Prime power:

120 / 240 V  
50 - 60 Hz

70 db ATTEN  
≈ 320 VOLTS/PK  
PRF ≈ 2 KHz