


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APD-2W-C-PN*

S.N.: *5591*

Date: *SEPT 28 1990*



- a) Output signal amplitude:
0 TO ± 10 AMPS
- b) Pulse width:
2 TO 50 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

50 db ATTEN

≈ 32 VOLTS/DIV

R_L = 5.0 Ω

≈ 6 AMP/DIV

PRF = 20 KHz