


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MM-2-C-P-M*

S.N.: *5603*

Date: *Jul 20 1990*



*40 db ATTEN*

*1.0 VOLTS/DIV*

*PRF ≈ 12 MHz*

- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*0.2 TO 2.0 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*

*..?*