


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MH-5-1-C-P*
S.N.: *5610*
Date: *SEPT 24 1990*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 30 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40dB ATTEN
10 VOLTS/DIV
PRF ≈ 1.0 MHz

120 / 240 V
50-60 Hz

[Signature]