


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-P*
S.N.: *5611*
Date: *AUG 23 1990*



50 db ATTEN
32 VOLTS/DIV
PRF = 50 KHz

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
2 TO 100 nSEC
- c) Rise time:
≤ 350 ps
- d) Fall time:
≤ 2 ns
- e) PRF:
MIN PW: 0 TO 100 Hz
MAX PW: 0 TO 20K Hz
- f) Jitter, stability: *(SEE FIG 142)*
OK
- g) Prime power:
+15 ± 1.0 VOLT
150 mA MAX

[Signature]