


PULSE GENERATOR

PERFORMANCE CHECK


Model: *APR-3-PW-C-P-CA-U1A*

S.N.: *5618*

Date: *OCT 5 1990*



①  $R_L = 50 \Omega$   
 $PRF = 2.0 \text{ KHz}$



②  $R_L = 50 \Omega$   
 $PRF = 2.0 \text{ KHz}$   
(MAX DUTY CYCLE)

- a) Output signal amplitude: *0 TO +200 V. TO*
- b) Pulse width: *50 ns*  
*0.1 TO 10 us*
- c) Rise time:  $\leq 10 \text{ ns}$
- d) Fall time:  $\leq 10 \text{ ns}$
- e) PRF: *0 TO 2 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 V*  
*50 - 60 Hz*
- h) DUTY CYCLE: *0 TO 2%*

*[Signature]*