


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MVI-V-HV3-C-PN-05*

S.N.: *5624*

Date: *SEPT 4 1990*



*(A) 60 dB ATTEN
100 VOLTS/DIV
PRF = 5.0 kHz*



*(B) AS (A) BUT 1.0 nSEC/DIV
(RISE TIME).*

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
4 TO 100 NS.
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120 / 240 V
5.0 - 60 kHz*