


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PR-A-1-PW-C-05-EA EN -PN*

S.N.: *5627*

Date: *OCT 25 1990*



- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: *TO 50 μs*
50 nSEC TO 1.0 μSEC
- c) Rise time: *0.5% DUTY CYCLE MAX*
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF: *0 TO 100 KHz*
(0.5% DUTY CYCLE MAX)
- f) Jitter, stability:
OK
- g) Prime power:

R_L = 50 Ω
PRF ≈ 10 KHz.

120 / 240 V
50-60 Hz