


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-B1-PS-W-T*

S.N.: *5627*

Date: *NOV 8 1990*



- a) Output signal amplitude:  
*0 TO +20.4028*
- b) Pulse width:  
*0.1 TO 5.0 μS*
- c) Rise time:  
*≤ 5 NS*
- d) Fall time:  
*≤ 5 NS*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*5.0 - 60 MHz*

*$R_L = 50 \Omega$*   
*PRF = 20 KHz*

*[Signature]*