


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MR-4A PW-PS-P*
S.N.: *5628*
Date: *NOV 8 1990*



- a) Output signal amplitude:
0 TO + 400 VOLTS
- b) Pulse width: *TO 50 ns*
0.1 TO 5.0 US
- c) Rise time:
≤ 20 NS
- d) Fall time:
≤ 20 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
30-60 W

R_i = 50 Ω
PRF = 1.0 KHz