


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PW-C-P*

S.N.: *5630*

Date: *SEPT 19 1990*



a) Output signal amplitude:
0 TO +200 VOLTS

b) Pulse width: *TO 50 n*
0.1 TO 100 uS.

c) Rise time:
≤ 10 NS


d) Fall time:
≤ 10 NS

e) PRF:
MAX PW: 0 TO 50 Hz

f) Jitter, stability:
PW ≤ 5 uS: 0 TO 1 kHz

g) Prime power: *OK*
120/240 V
50-60 Hz

Ⓐ *50 VOLTS/DIV*
200 NS/DIV
RL = 50 n



Ⓑ *AS Ⓐ BUT*
20 uS/DIV.