



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-U-C-N
S.N.: 5633
Date: OCT 9 1990



(A) 50 dB ATTEN
32 VOLTS/DIV
PRF = 20 KHz



(B) 50 dB ATTEN
32 VOLTS/DIV
PRF = 20 KHz
(RISE TIME)

- a) Output signal amplitude:
0 TO -50 VOLTS.
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]