


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-64-P
S.N.: 5640
Date: JAN 4 1991



60 dB ATTEN
≈ 100 VOLTS / DIV
PRF = 70 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: (70 50 ns)
1 TO 5 NS
- c) Rise time:
≤ 0.4 NS
- d) Fall time:
≤ 0.4 NS
- e) PRF: 0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VOLTS
250 mA (MAX)