


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-C-N-OT-FOICA*

S.N.: *5642*

Date: *OCT 19 1990*



*40 dB ATTEN*  
*5.0 VTS DIV*  
*~ 220 MHz*

- a) Output signal amplitude:  
*0 TO -5.0 VTS*
- b) Pulse width:  
*0.2 TO 1.0 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 100 PS*
- e) PRF:  
*50 TO 250 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*5.0 to 60 Hz*