


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMMN-1-PS-P*


S.N.: *8644*

Date: *OCT 23 1990*



- a) Output signal amplitude:
0 TO +5 VOLTS.
- b) Pulse width:
0.2 TO 5.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 MHz

ⓐ *40db ATTEN*
5 VOLTS/DIV
PRF ≈ 50 MHz



ⓑ *AS ⓐ BUT*
1 NS/DIV

[Signature]