


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVN-1-P

S.N.: 5645

Date: OCT 23 1990



40 db ATTEN  
5.00V/DIV

a) Output signal amplitude:

0 TO +5 VOLTS

b) Pulse width:

0.2 TO 1.0 NS

c) Rise time:

$\leq 100$  PS

d) Fall time:

$\leq 100$  PS

e) PRF:

50 TO 250 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 VOLT

500 mA (MAX)