


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MU-PN-55B*

S.N.: *5650*

Date: *NOV 1 1990*



- a) Output signal amplitude:
P_{out} : \approx 100 VOLTS
- b) Pulse width:
P_{out} : \leq 100 VOLTS
- c) Rise time: *0.6 TO 2.0 NS*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLTS*
200 mA (MAX)

P_{out}
60 db ATTEN
100 VOLTS/DIV
PRF = 70 KHz