


PULSE GENERATOR
PERFORMANCE CHECK

Model: AYR-3-pw-PST
S.N.: 5659
Date: OCT 30 1990



- a) Output signal amplitude:
0 TO +200 μ V
- b) Pulse width:
0.1 TO 100 μ SEC.
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
MAX PW: 0 TO 50 MHz
PW ≤ 10 μ SEC: 0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

$R_L = 50 \Omega$
PRF = 500 MHz

120 / 240 V
50 - 60 MHz