


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PR-A-2-PW-C-P*

S.N.: *5663*

Date: *NOV 12 1990*



*50 dB ATTEN*

*≈ 3.2 VOLTS / DIV*

*PRF = 1.0 MHz*

- a) Output signal amplitude:  
*0 TO +50.0 VOLTS*
  - b) Pulse width:  
*20 TO 200 NS*
  - c) Rise time:  
*(20% MAX DUTY CYCLE)*  
*≤ 5 NS*
  - d) Fall time:  
*≤ 5 NS*
  - e) PRF:  
*0 TO 10 MHz*
  - f) Jitter, stability:  
*(20% MAX DUTY CYCLE)*  
*OK*
  - g) Prime power:  
*120 / 240 VOLTS*  
*50-60 Hz*
- A*