


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMR-B3.C-W-P*

S.N.: *5664*

Date: *OCT 31 1990*



- a) Output signal amplitude:  
*0 TO 7700 VOLTS*
- b) Pulse width:  
*0.1 TO 5.0 US*
- c) Rise time:  
*2.5 NS*
- d) Fall time:  
*2.5 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*3.0 60 Hz*

*Av = 50V*  
*PRF = 5 KHz*

*[Signature]*