


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PNH-C-P*  
S.N.: *5665*  
Date: *NOV 14 1990*



*40 db ATTEN*  
*10 VOLT DIV*  
*PRF  $\approx$  1.0 MHz*

- a) Output signal amplitude:  
*0 TO 730 VOLTS*
- b) Pulse width:  
 *$\leq$  1.0 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240V*  
*50-60 Hz*  
*[Signature]*