


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PK-A.P*  
S.N.: *5683*  
Date: *NOV 5 1990*



- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*1 TO 50 NS*
- c) Rise time:  
*< 100 PS*
- d) Fall time:  
*< 2 NS*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN*  
*10 VOLT/DIV*  
*PRF = 50 KHz*

*+15 VOLTS*  
*350 mA (MAX)*  
*[Signature]*