


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-S-N*
S.N.: *5687*
Date: *NOV 5 1990*



- a) Output signal amplitude:
0 TO -15 VOLTS
- b) Pulse width:
± 500 PS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV
PRF = 700 KHz

+15 VOLTS
150 mA
MIAT

[Signature]