


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-64-PN-ATT-C

S.N.: 5701

Date: NOV 23 1990



a) Output signal amplitude:
 $P_{out} = 0 \text{ TO } +200 \text{ V}$) $R_L \geq 100\Omega$
 $V_{out} = 0 \text{ TO } -200 \text{ V}$
 b) Pulse width:
 0.5 μ S TO 5.0 MS

c) Rise time:
 $\leq 10 \text{ NS}$


d) Fall time:
 $\leq 10 \text{ NS}$

e) PRF: 100 Hz TO 1.0 MHz

f) Jitter, stability:
 OK

g) Prime power:
 120/240 V
 50 - 60 Hz

(A) PRF = 1.0 MHz.
 50 VOLTS/DIV



(B) PRF = 10 KHz.
 50 VOLTS/DIV