


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-7C-E-N-M-AT*


S.N.: *5702*

Date: *DEC 14 1990*



10 uS/DIV
Rc = 1.0 Ω
PRF = 50 kHz

- a) Output signal amplitude:
0 TO -30 AMP TO
- b) Pulse width: *0 TO -300 uS*
0.2 TO 50 uS
- c) Rise time:
≤ 50 NS
- d) Fall time:
≤ 50 NS
- e) PRF:
MAX PW : 0 TO 50 kHz
MIN PW : 0 TO 1 kHz.
- f) Jitter, stability:
OK



0.5 uS/DIV
Rc = 1.0 Ω
PRF = 1.0 kHz

- g) Prime power:
120 / 240 V
50-60 Hz
- [Signature]*