


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PVI-V-14V3-PS-PN*

S.N.: *5706*

Date: *MAR 17 91*



Ⓐ 60db ATTEN
100 V/DIV
PRF = 50 KHz



Ⓑ AS Ⓐ BUT
1.0 NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 100 VOLTS
 - b) Pulse width:
5 TO 100 NS
 - c) Rise time:
≤ 200 PS
 - d) Fall time:
≤ 3 NS
 - e) PRF:
0 TO 5 KHz
 - f) Jitter, stability:
OK
 - g) Prime power:
120/240 V
50-60 Hz
- [Signature]*