


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVG-4A-PS-PN

S.N.: 5707

Date: JAN 17 91



- a) Output signal amplitude:  
0 TO  $\pm 650$  VOLTS
- b) Pulse width:  
 $\leq 5$  NS.
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 20 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

6.0 db ATTEN  
100 VOLTS/DIV  
PRF = 7.0 KHz

120/200 V  
50-60 Hz  
*[Signature]*