


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMP-M-2-C-05-P-PN*

S.N.: *5715*

Date: *NOV 16 1990*



Ⓐ *40 db ATTEN*
10 VOLTS/DIV
PRF = 50 KHz



Ⓑ *AS Ⓐ BUT 1 NS/DIV*
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 10 V
- b) Pulse width:
5 TO 50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

[Signature]