


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVO-1-C-P-PN-MD-2

S.N.: 5717

Date: JAN 16 1991



P_{out} C_{ERT} ≈ 500 pfd
≈ 60 AMP/DIV



P_{out} C_{ERT} ≈ 0 pfd.
≈ 60 AMP/DIV



N_{out} C_{ERT} ≈ 10,000 pfd
≈ 60 AMP/DIV

- a) Output signal amplitude:
±25 TO ±100 AMPs
- b) Pulse width:
10 TO 50 NS
- c) Rise time: (HALF POWER)
NA
- d) Fall time:
NA
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz.